

## Endotoxin Quantitative Kinetic Turbidimetric Assay Report of Analysis

### Product Information

**Client:** LARP LLC  
**Product Tested:** GLOW “Beverly Hills”  
**Lot Number:** 2KGLW08252025  
**Vials Tested:** 1  
**Test Performed:** PYROSTAR ES-F/Plate Kinetic Turbidimetric Assay  
**Test Date:** 03/30/2026

### Reagent Information

**Lysate Product #:** WPEPK4-50015  
**Lysate Lot #:** 122412A  
**Lysate Exp:** 12/01/2026  
**Lysate Sensitivity:** 0.01 – 10 EU/mL  
**Control Standard Endotoxin (CSE) Product #:** WPEPK4-50015  
**CSE Lot #:** G-07USA  
**CSE Exp:** 09/01/2029


### KTA Test Results

| Solution                       | Endotoxin Concentration/Added to | Concentration (EU/ml) / % Recovery                             |
|--------------------------------|----------------------------------|--|
| Sample/Product Solution        | None (EU/mL)                     | <b>A: &lt;0.079, B: &lt;0.079</b><br><b>Average: &lt;0.079</b> |
| Positive Product Control (PPC) | 1.0 EU/mL in Product             | <b>A: 1.091, B: 1.122</b><br><b>Percent Recovery: 110.617%</b> |
| Negative Control               | None (EU/mL)                     | <b>A: &lt;0.079, B: &lt;0.079</b>                              |

**Interpretation:** This test is considered valid when: the positive product control % recovery is between 50% and 200%, and both negative control concentrations do not exceed the limit of the blank value required by the lysate reagent employed (0.079 EU/ml). When sample concentrations match the negative control, samples are considered free of endotoxin.

Sample resuspended in 2 ml endotoxin free water for testing. **Average EU/ml in test solution: <0.079**

**TOTAL Average Endotoxin EU in Sample: NEGATIVE**

Analyst Signature:  \_\_\_\_\_